

**Scope of Accreditation  
For  
Canada West Torque Tools  
Division of McCANN Equipment LTD**

#7-7533-135<sup>th</sup> Street  
Surrey, British Columbia, V3W 0N6 Canada  
Daniel M. McCann  
604-596-4077

In recognition of a successful assessment to ISO/IEC 17025:2005 and ANSI Z540-1, accreditation is granted to **Canada West Torque Tools, Division of McCANN Equipment LTD** to perform the following **Calibrations**:

Accreditation granted through: **June 29, 2010**

**Calibration**

<b>Calibration Parameter/Equipment<sup>1</sup></b>	<b>Range</b>	<b>Best Measurement Capability (+/-)<sup>2</sup></b>	<b>Remarks</b>
Hand Torque Wrenches and Torque Limiting Screwdrivers	4 ozf·in to 2000 lbf·ft 2.8 cNm to 2712 Nm	0.48% of applied load	ISO 6789
Pneumatic Torque Wrenches	0.4 lbf·in to 5000 lbf·ft 0.04 cNm to 6780 Nm	1.52% of applied load	Electronic transducer and display unit per McCANN procedures
Hydraulic Torque Wrenches	127 lbf·ft to 5000 lbf·ft 172 Nm to 6780 Nm	1.17% of applied load	Electronic transducer and display unit per McCANN procedures
Electronic Torque Wrenches	4 ozf·in to 2000 lbf·ft 2.8 cNm to 2712 Nm	0.6% of applied load	Electronic transducer and display unit per McCANN procedures
Hand Torque Multipliers	127 lbf·ft to 5000 lbf·ft 172 Nm to 6780 Nm	0.59% of applied load	Electronic transducer and display unit per McCANN procedures

Notes:

- 1) Laboratory offers calibration services at the laboratory's own facilities and at the client or other agreed upon facilities.
- 2) Best uncertainties represent expanded uncertainties at approximately the 95% confidence level using a coverage factor of k=2.

Approved by: 

Date: June 15, 2007

R. Douglas Leonard Jr.  
Chief Technical Officer

Issued: 6/29/04 Revised: 5/3/05 Revised: 3/28/06 Revised: 6/15/07

Form 28.8—Rev 6 June 10, 2003